



CERTIFICATION by CUSTODIAN of RECORDS  
Tucson Police Department Crime Laboratory

104

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 5000 Serial Number 04-003336 (#15) maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 7/1/05 Cal 1 — AC Cal 2 — AC

Date: — Cal 1 — AC Cal 2 — AC

Date: — Cal 1 — AC Cal 2 — AC

Between the dates of — and — the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 7/1/05.

The instrument was returned to service on —.


Irene Doran

Irene Doran

7/18/05

ADPS QAS Permit # QAS 29586

SUBSCRIBED AND SWORN to before me this 18, day of July, 2005



Kathy A Kinney

Notary Public

\*Includes documents specified in A.A.C. R9-14-404.A.6



# MEMORANDUM

---

**DATE:** July 13, 2005

**TO:** Applied Electronics  
448 Grand Avenue  
Eagle, CO 81631

**FROM:** Irene Doran *ld*  
Tucson Police Department  
Crime Laboratory  
270 S. Stone Avenue  
Tucson, AZ 85701-1917

**SUBJECT:** Intoxilyzer 5000 Maintenance and Repair

Enclosed is Intoxilyzer 64-003336 (#15). The instrument was brought to the lab after an officer reported that the instrument was not working and that the pump was not turning on. When I turned the instrument on at the lab, the instrument had a Processor Error 1.

Please repair as needed.



TUCSON POLICE DEPARTMENT  
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

SM

Date: 7-7-05

QAS: Irene Doran #46331

Intoxilyzer Serial Number: (#15) 64-003336

Reason For Service: Officer <sup>Dugan</sup> reported pump not coming on. 7/11/05  
At lab processor Error 1.

Date Out of Service: 7/11/05

Calibration Check completed prior to removal from service

Date: NA

SQAP completed prior to removal from service

Date: NA

Corrective Action:

☒ Sent to Applied Electronics



# MEMORANDUM

---

**DATE:** July 12, 2005

**TO:** Whom it may concern

**FROM:** Irene Doran  
Intoxilyzer 5000 QAS  
Tucson Police Department  
Crime Lab  
270 S. Stone Avenue  
Tucson, AZ 85701-1917  
(520) 791-4494

**SUBJECT:** Intoxilyzer Out of Service

Intoxilyzer 5000, serial number **64-003336 (#15)**, was taken out of service for repair on **7/1/05**.

I cannot testify that it was working properly and accurately between the dates of  
**June 7, 2005 and July 1, 2005.**

This instrument was located at **TRAFFIC**.